



# Certificate of Conformance

ISO 9001 Quality Management certified by  
BSI under certificate number FM595556

|             |              |                  |   |          |        |
|-------------|--------------|------------------|---|----------|--------|
| Part Number | FFNN2545-100 | Part Description | Cronus 0.45 µm Nylon 25 mm Syringe Filter (Green) (Jar/100) |          |        |
| Lot Number  | XXXXXX       | ID Number        | XXXXXX  | Quantity | XXXXXX |

Quality Lab Accessories, LLC, certifies, warrants and represents that the filters described herein conform to the specifications set forth in U.S. Pharmacopoeia, General Chapter, <1092> Dissolution Procedure: Development and validation/ Filters) as last revised, as well as ASTM E 2503-13 and were designed, manufactured or fabricated within strict compliance of these same standards.

| Test                             | Results  |
|----------------------------------|--|
| Pore Size                        | 0.45 µm  |
| Diameter                         | 25 mm  |
| Leak Testing:                    | > 75 psi   |
| Extractables                     | Within Limits as Specified in USP                      |
| Integrity Testing:               | Each Unit is tested for membrane and housing integrity |
| Filtration Area:                 | 4.91 cm <sup>2</sup>                                   |
| Minimum Effective Burst Pressure | 14 Kg/cm <sup>2</sup>                                  |



|                        |        |
|------------------------|--------|
| Certified by Inspector | Date   |
| XXXXXXXXXX             | XXXXXX |